



## Quality Certificate

### VWR\* NITRILE EXAM GLOVES

This document certifies that the below product meets the following criteria:

North American Catalog No:	82026-426
Lot Number:	234R11A303
Description:	VWR GLOVE NITRILE PF M PK100
Manufacturing Date:	January 2011
Product Shelf life:	December 2016

**Material:** Nitrile Butadiene Rubber, 100% Latex Free

**QC Testing:** Representative production samples are collected and inspected in accordance with current applicable product specification:

Meet or exceed ASTM D 6319-00a(Water test & Dimension)  
Meet or exceed EN 455-1&2  
ISO 2859 (Sampling table)

**Accuracy:**

Water tight test:	Inspection level=G1; AQL 1.5
Visual Inspection:	Inspection level=G1; AQL 1.5(critical defects) AQL 2.5(major defects) AQL 4.0(minor defects)
Dimensional:	Inspection level=S2; AQL 4.0
Physical Property:	Inspection level=S2; AQL 4.0
Powder Content:	Inspection level=5pieces; Average<=2mg/glove
Moisture Content:	Inspection level=5pieces; Average<=1%
Glove Count:	Inspection level=5dispensers;Tolerance+/-2 pieces

**Sterility:** N/A

**Non-pyrogenic:** N/A

**Rnase/Dnase Free:** N/A

**Regulatory:** FDA 510K, ISO 9001, ISO 13485, CMDCAS Quality System, GMP Certificate

Signed:

Ken Crossley  
Manager  
Quality Assurance

Date of Issue: Jan 31, 2011