



Quality Certificate

VWR* GLOVES

This document certifies that the below product meets the following criteria:

| | |
|-----------------------------------|------------------------------|
| North American Catalog No: | 82026-426 |
| Lot Number: | PT234R11S004 |
| Description: | VWR GLOVE NITRILE PF M PK100 |
| Manufacturing Date: | Dec 2011 |
| Product shelf Life: | Nov 2016 |

Material: Nitrile Butadiene Rubber, 100% latex Free

QC Testing: Representative production samples are collected and inspected in accordance with current applicable product specifications.

Meet or exceed ASTM D 6319-00a (Water Test & Dimension)
Meet or exceed EN 455-1&2
ISO 2859 (sampling table)

Accuracy:

| | |
|---------------------------|--|
| Water tight test: | Inspection level = G1; AQL1.5 |
| Visual inspection: | Inspection level = G1; AQL1.5 (critical defects) AQL2.5 (major defects) AQL4.0 (minor defects) |
| Dimensional: | Inspection level = S2; AQL4.0 |
| Physical Property: | Inspection level = S2; AQL4.0 |
| Powder Content: | Inspection level = 5pieces; Average ≤ 2 mg/glove |
| Moisture Content: | Inspection level = 5pieces; Average $\leq 1\%$ |
| Glove Count: | Inspection level = 5 dispensers; Tolerance 100+/-2pieces |

Sterility: N/A

Non-Pyrogenic: N/A

Rnase/Dnase Free: N/A

Regulatory: FDA510K, ISO 9001, ISO 13485, CMDCAS Quality System, GMP Certificate

Signed:

Ken Crossley
Manager
Quality Assurance

Date of Issue: December 28, 2011