



Quality Certificate

VWR* SPECTROPHOTOMETER CELL

This document certifies that the below product meets the following criteria:

North American Catalog No:	414004-062
Batch Number:	2013011223
Description:	VWR CELL QUARTZ FLAT 10MM PK2

Material : Quartz

Quantity : 5CS

INTERIOR		EXTERIOR			TRANSMISSION
Length	Width	Length	Width	Height	200nm>80%
10±0.05	10±0.2	12.5±0.1	12.5±0.1	45±0.5	

Path Length : 10mm

Signed:

Ken Crossley
Manager
Quality Assurance

Date of Issue: January 12, 2013