



Quality Certificate

VWR* GLOVES

This document certifies that the below product meets the following criteria:

Product Number	82026-424
Lot Number	PT234R12L202
Description	VWR GLOVE NITRILE PF S PK100
Manufacturing Date:	September 2012
Product Shelf Life:	August 2017

Material: Nitrile Butadiene Rubber , 100% Latex Free

QC Testing: Representative production samples are collected and inspected in accordance with current applicable product specification.

Meet or exceed ASTM D 6319-00a (Water Test & Dimension)

Meet or exceed EN 455-1&2

ISO 2859 (sampling table)

Accuracy: Water tight test: Inspection level=G1;AQL 1.5
Visual inspection: Inspection Level=G1;AQL 1.5 (criticaldefects)
AQL 2.5 (major defects)
AQL 4.0 (minor defects)

Dimensional: Inspection Level= S2;AQL 4.0
Physical property: Inspection Level=S2; AQL 4.0
Powder Content: Inspection Level=5pieces ; Average ≤ 2 mg/glove
Moisture Content: Inspection level=5pieces; Average $\leq 1\%$
Glove Content: Inspection level=5dispensers;
Tolerance 100+/-2 pieces

Sterility: N/A

Non-Pyrogenic: N/A

Rnase/Dnase Free: N/A

Regulatory: FDA 510K,ISO 9001,ISO 13485 under CMDCAS Quality System, GMP certificate

Signed:

Ken Crossley
Manager
Quality Assurance

Date of Issue: September 16, 2012