



## Quality Certificate

### VWR\* GLOVES

This document certifies that the below product meets the following criteria:

<b>Product Number:</b>	82026-428
<b>Lot Number:</b>	234R10J013(L)
<b>Description:</b>	VWR GLOVE NITRILE PF L PK100

**Material:** Nitrile Butadiene Rubber

**QC Testing:** Representative production samples are collected and inspected in accordance with current applicable product specification.

Meet or exceed ASTM D 6319-00a (Water Test & Dimension)  
Meet or exceed EN 455-1&2  
ISO 2859 (sampling table)

**Accuracy:**

Water tight test:	Inspection level=G1;AQL 1.5
Visual inspection:	Inspection Level=G1;AQL 1.5 (criticaldefects) AQL 2.5 (major defects) AQL 4.0 (minor defects)
Dimensional:	Inspection Level= S2;AQL 4.0
Physical property:	Inspection Level=S2; AQL 4.0
Powder Content:	Inspection Level=5pieces ; Average $\leq 2$ mg/glove
Moisture Content:	Inspection level=5pieces; Average $\leq 1\%$
Glove Content:	Inspection level=5dispensers; Tolerance 100+/-2 pieces

**Sterility:** N/A

**Non-Pyrogenic:** N/A

**Rnase/Dnase Free:** N/A

**Regulatory:** FDA 510K,ISO 9001,ISO 13485 under CMDCAS, GMP certificate

Signed:

Ken Crossley  
Manager  
Quality Assurance

Date of Issue: Jul 09, 2010