

CERTIFICATE OF COMPLIANCE

Page: 1 of 2

Produc	t I	Name
Catalo	g I	Number
Batch	Nu	mber

: Syringe 10ml LL : 309604 : 3171430

Manufacture Date : 2013/08/07

REGULATORY COMPLIANCE AND QUALITY SYSTEM

BD Products comply with the regulatory requirements of the region in which these are sold and manufactured.

BD Products sold in the US comply with the current FDA Quality System Regulation 21CFR820. Medical devices are listed with FDA per 21CFR807. Manufacturing sites are registered with FDA per 21CFR807. The devices satisfy FDA pre-market notification requirements per 21CFR 807.

STERILITY

All products which are labeled as "sterile" and released for sale by BD are certified to be sterile as long as the package is unopened and undamaged.

This product is primarily sterilized via E-beam. Sterilization cycle development/validation is performed to 10-6 SAL in accordance with current ISO 11137 guidelines.

BIOCOMPATABILITY

This product has been evaluated in accordance with ISO 10993 "Biological Evaluation of Medical Devices", and complies with all relevant sections.

PYROGENICITY

All products which are labeled as non-pyrogenic and released for sale by BD have been tested per United States Pharmacopeia (USP) chapter 85 - Bacterial Endotoxins Test and meets limits as specified in chapter 161- Transfusion and Infusion Assemblies and Similar Medical Devices.

QUALITY CONTROL TESTING AND RELEASE

Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and signed off by qualified personnel for product



CERTIFICATE OF COMPLIANCE

Page: 2 of 2

٦

Product Name: Syringe 10ml LLCatalog Number: 309604Manufacture Date : 2013/08/07Batch Number: 3171430		
release. The released devices meet applicable BD product specifications.		
PRODUCT SPECIFIC SPECIFICATIONS / STANDARDS		
This product complies with the following BD Specification: SP100001		
BD MANUFACTURING SITE		
BD Manufacturing site: BD Medical, Route 7 and Grace Way, Canaan, CT 06018, USA		
Primary Sterilization site: BD Medical, Route 7 and Grace Way, Canaan, CT 06018, USA		
Legal Manufacturer: Becton, Dickinson and Company, 1 Becton Drive, Franklin Lakes, NJ 07417, USA		

Mary Rivers

Mary Rivers Quality Manager, BD Canaan