

**Corning Incorporated  
Life Sciences**

**Registered  
ISO 9001:2008**

# Certificate of Compliance

**Catalog Number:** 2116-50BO  
**Lot Number:** 12812065  
**Product Description:** PYREX® Buret, Body Only, Economy Grade, 50 mL

**Lot Number Designation:**

This production lot has been uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

**Component Materials:**

Type I Class A - Low-Expansion Borosilicate Glass, per ASTM E438 (Standard Specification for Glasses in Laboratory Apparatus)

**Attributes:**

Capacity - 50 mL ±0.10 mL at any mark

This product has been sampled in accordance with Standard Operating Procedure and has been released by Quality Assurance for the following characteristics:

<b>Test/Procedure:</b>	<b>Results:</b>
Visual Attributes	Pass
Dimensional	Pass
Packaging	Pass
Capacity	Pass

**Revision Date:**

18 JUN 2013



Frigyes Szendro Kassa  
Quality Manager

Rev No: 1