



## Quality Certificate

### VWR\* GLOVES

This document certifies that the below product meets the following criteria:

<b>Product Number</b>	82026-430
<b>Lot Number</b>	PT234R13E107
<b>Description</b>	VWR GLOVE NITRILE PF XL PK100
<b>Manufacturing Date:</b>	May 2013
<b>Product Shelf Life:</b>	April 2018

**Material:** Nitrile BUTADIENE Rubber, 100% Latex Free

**QC Testing:** Representative production samples are collected and inspected in accordance with current applicable product specification.

Meet or exceed ASTM D 6319-00a (Water Test & Dimension)

Meet or exceed EN 455-1&2

ISO 2859 (sampling table)

**Accuracy:**

Water tight test:	Inspection level=G1; AQL 1.5
Visual inspection:	Inspection Level=G1; AQL 1.5 (critical defects) AQL 2.5 (major defects) AQL 4.0 (minor defects)
Dimensional:	Inspection Level= S2; AQL 4.0
Physical property:	Inspection Level=S2; AQL 4.0
Powder Content:	Inspection Level=5pieces ; Average ≤2mg/glove
Moisture Content:	Inspection level=5pieces; Average≤1%
Glove Count:	Inspection level=5dispensers; Tolerance 100+/-2 pieces

**Sterility:** N/A

**Non-Pyrogenic:** N/A

**Rnase/Dnase Free:** N/A

**Regulatory:** FDA 510K,ISO 9001,ISO 13485 under CMDCAS Quality System, GMP certificate

**Signed:**

Ken Crossley  
Manager  
Quality Assurance

**Date:** March 29 , 2013