## Corning Incorporated Life Sciences

Registered ISO 9001:2008

## Certificate of Compliance

Catalog Number: 9800-13

Lot Number: 17514016

Product Description: PYREX® Tube, Test, Beaded Rim, 13 x 100 mm

Lot Number Designation:

This production lot has been uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

**Component Materials:** 

Type I Class A - Low-Expansion Borosilicate Glass, per ASTM E438 (Standard Specification for Glasses in Laboratory Apparatus)

Attributes:

Capacity - 9 mL

TSE/BSE:

Materials of animal origin are not used in the manufacture of this product.

This product has been sampled in accordance with Standard Operating Procedure and has been released by Quality Assurance for the following characteristics:

Test/Procedure:Results:Visual AttributesPassDimensionalPassPackagingPass

Revision Date: 11 MAR 2013

Racelis Acosta Reynoso Plant Quality Manager

Rev No: 1