



Quality Certificate

VWR* NITRILE GLOVE

This document certifies that the 4 mil Powder Free Nitrile textured gloves meet the following criteria:

North American Catalog No:	82026-424
Lot Number:	PT234R12L302
Description:	VWR GLOVE NITRILE PF XS PK100
Manufacturing Date:	September,, 2012
Product shelf Life:	August, 2017

Material: Nitrile BUTADIENE Rubber, 100% latex Free

QC Testing: Representative production samples are collected and inspected in Accordance with current applicable product specifications:

Meet or exceed ASTM D 6319-00a (Water Test & Dimension)
Meet or exceed EN 455-1&2
ISO 2859 (Sampling table)

Accuracy:

Water tight test:	Inspection level = G1; AQL1.5
Visual inspection:	Inspection level = G1; AQL1.5 (critical defects) AQL2.5 (major defects) AQL4.0 (minor defects)
Dimensional:	Inspection level = S2; AQL4.0
Physical Property:	Inspection level = S2; AQL4.0
Powder Content:	Inspection level = 5pieces;Average ≤ 2 mg/glove
Moisture Content:	Inspection level = 5pieces; Average $\leq 1\%$
Glove Count:	Inspection level = 5 dispensers; Tolerance 100+/-2pieces

Sterility: N/A

Non-Pyrogenic: N/A

Rnase/Dnase Free: N/A

Regulatory: FDA510K, ISO 9001, ISO 13485, CMDCAS Quality System, GMP Certificate

Signed:

Ken Crossley
Manager
Quality Assurance

Date of Issue: September 18, 2012