



# NIPRO MEDICAL INDUSTRIES LTD.

2-19-64, Matsubara, Tatebayashi-shi, Gunma, 374-8518 Japan  
TEL:0276-72-5117 FAX:0276-74-6723

TO : Becton Dickinson Vacutainer System  
Attention : BD PAS Quality Assurance Department

Date : Apr 11,2014

## Certificate of Conformance

Purchase Order No. : 4503180328  
Reorder No. : 367296  
Lot No. : 4C3181  
Quantity : 132000 pcs  
Date of Manufacture : Apr 3,2014  
Expiry : 2017-03  
Specification : VS59838  
Revision : 19

This is to certify that the item identified herein has been tested in accordance with the specification of VS59838 and meets its requirements. It is further certified that no material/process substitutions were made for any components specified in VS59838.

Test	Sample	AQL	Judgement
Biological Indicator	22 BI/Chamber	0	Passed
Bacterial Endotoxins	10 Set	< 0.5EU/mL	Passed

### Method

#### Biological Indicator

U.S. Pharmacopeia 36:2013.

<71>Sterility Test

#### Bacterial Endotoxins

U.S. Pharmacopeia 36:2013.

<85>Bacterial Endotoxins Test

<161>Transfusion and Infusion Assemblies

Monitoring	Tolerance	Judgement
Process parameters	Within limit	Passed

Method :ISO11135-1:2007.

Authorized Signature :

Minoru Oishi  
Quality Control Manager  
Nipro Medical Industries Ltd.