

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Certificate of Compliance

Catalog Number: 2116-50BO
Lot Number: 28014048
Product Description: PYREX® Buret, Body Only, Economy Grade, 50 mL

Lot Number Designation:

This production lot has been uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Component Materials:

Type I Class A - Low-Expansion Borosilicate Glass, per ASTM E438 (Standard Specification for Glasses in Laboratory Apparatus)

Attributes:

Capacity - 50 mL ±0.10 mL at any mark

This product has been sampled in accordance with Standard Operating Procedure and has been released by Quality Assurance for the following characteristics:

Test/Procedure:	Results:
Visual Attributes	Pass
Dimensional	Pass
Packaging	Pass
Capacity	Pass

Revision Date:

18 JUN 2013



Racelis Acosta Reynoso
Plant Quality Manager

Rev No: 1