

## **CERTIFICATE OF CALIBRATION**

**Model 402KL**

**Capacity: 500 lb. x ¼ lb. / 200 kg. x 100 g  
(With included 150 lb. / 40 kg counterweight)**

This is to certify that all the products distributed by Pelstar LLC are calibrated at the time of manufacturing to tolerances in the specifications. This does not certify that the products have not been altered by tampering or damaged by misuse or mishandling. This is not meant to supersede individual or internal calibration programs.

### **SECTION 1 Calibration Method**

Scales are calibrated by direct application of certified dead weight to the weighing platform. Calibration procedure may only be performed by approved operators. Pelstar LLC manufacturing procedures requires verification of conformance to product specifications by production personnel at the time of manufacturing. Product release is contingent on acceptance by Inspection personnel to all specification in place.

### **SECTION 2 Calibration Procedures**

Only weights certified and traceable to national standards are use for calibration/verification procedures