



Becton Dickinson Caribe LTD.  
 PO Box 372860  
 Cayey PR 00737-2860 PR

# Certificate of Analysis

Product Name : DISC AMPICILLIN AM-10 10 EA  
 Catalog Number : 231264      Manufacture Date : 2011/07/21  
 Batch Number : 1199483  
 Expiration Date : 2013/07/31

Characteristic	Unit	Value	LowLimit	HighLimit
Micro Assay Range:		10.94		
Micro Assay Range:		11.06		
Micro Assay Range:		11.88		
Micro Assay Range:		11.50		
Micro Assay Potenc		11.00		
Uniformity		0.8	-	2.5
Uniformity		1.0	-	2.5
Uniformity		0.6	-	2.5
Uniformity		0.6	-	2.5

The Batch Number on this certificate is synonymous with the Lot Number shown on the product label.

BD Diagnostic Systems (BDDS) is an ISO 13485:2003 and ISO 9001:2008 Registered facility. BDDS products are manufactured in facilities registered with the United States Food and Drug Administration (FDA), and are regulated by the FDA's Quality System Regulations (QSRs). This product met BDDS stringent quality standards at time of batch/lot release. Any test results reported on this certificate were obtained at time of release.

**John Gerlich**  
 Vice President,  
 Quality Management and  
 Regulatory Compliance  
 Signature Date: 2011/08/17