



Quality Certificate

VWR® SPECTROSIL SPECTROPHOTOMETER CELLS

This document certifies that the below product meets the following criteria:

North American Catalog No:	414004-062
Lot Number:	2010071721
Description:	VWR CELL QUARTZ FLAT 10MM PK2

Material : Quartz

INTERIOR		EXTERIOR			TRANSMISSION
Length	Width	Length	Width	Height	200nm>80%
10±0.05	10±0.3	12.5±0.1	12.5±0.1	45±0.3	

Path Length : 10mm

Signed:

Ken Crossley
Manager
Quality Assurance

Date: July 17, 2010