



Quality Certificate

VWR® NITRILE EXAMINATION GLOVE

As per the manufacturer, the below product meets the following criteria:

| | |
|------------------------------------|---|
| North American Catalog No: | 82026-426 |
| Lot No: | PT234R15H104 |
| Description: | VWR GLOVE NITRILE PF M PK100 |
| Manufacturing Date: | June 2015 |
| Product Shelf Life : | May 2020 |
| Country Of Origin : | Indonesia |
| Quality System Compliance : | FDA 510K, ISO 9001, ISO 13485, CMDCAS Quality System ,GMP Certificate |

Material: Nitrile Butadiene Rubber, Not made with natural rubber latex
QC Testing: Representative production samples are collected and inspected in accordance with current applicable Product specification:

- Meet or exceed ASTM D 6319-00a (Water test & Dimension)
- Meet or exceed EN 455 -1&2
- ISO 2859 (Sampling table)

| | | | |
|------------------|--------------------|----------------------------------|--|
| Accuracy: | Water tight test: | Inspection level = G1; | AQL 1.5 |
| | Visual inspection: | Inspection level = G1; | AQL 1.5 (critical defects) AQL 2.5 (major defects) AQL 4.0 (minor defects) |
| | Dimensional: | Inspection level = S2; | AQL 4.0 |
| | Physical Property: | Inspection level = S2; | AQL 4.0 |
| | Powder Content: | Inspection level = 5 pieces; | Average ≤ 2 mg/glove |
| | Moisture Content: | Inspection level = 5 pieces; | Average ≤ 1% |
| | Glove count: | Inspection level = 5 dispensers; | Tolerance 100+/-2 pieces |

Sterility: N/A
Non- Pyrogenic: N/A
Rnase/ Dnase Free: N/A
Cytotoxicity: N/A
BSE/ TSE Statement: N/A

Signed:



Ken Crossley
 Manager
 Quality Assurance

Date: June 28, 2015