


Date of Certificate: 08/11/2015
Product Number: 5000-0020
Product Lot Number: 1151775
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
1551427	07/31/2015	156	CS	20.5	25.6
1551460	07/31/2015	156	CS	20.1	25.6
1553096	07/31/2015	156	CS	20.2	25.3
1551503	07/31/2015	156	CS	20.4	25.5
1559446	08/05/2015	78	CS	20.1	25.9



Barry Srolis
Sr. Quality Manager