

## Quality Certificate

### VWR® NITRILE EXAMINATION GLOVE

As per the manufacturer, the below product meets the following criteria:

|                                    |                                                                       |
|------------------------------------|-----------------------------------------------------------------------|
| <b>North American Catalog No :</b> | 82026-424                                                             |
| <b>Lot No :</b>                    | PT234R15H302                                                          |
| <b>Description :</b>               | VWR GLOVE NITRILE PF S PK100                                          |
| <b>Date of Manufacturing :</b>     | June 2015                                                             |
| <b>Product Shelf Life :</b>        | May 2020                                                              |
| <b>Country of Origin :</b>         | Indonesia                                                             |
| <b>Quality System Compliance :</b> | FDA 510K, ISO 9001, ISO 13485, CMDCAS Quality System, GMP Certificate |

**QC Testing:** Representative production samples are collected and inspected in accordance with current applicable Product specification:

- Meet or exceed ASTM D 6319-00a (Water test & Dimension)
- Meet or exceed EN 455 -1&2
- ISO 2859 (Sampling table)

|                  |                    |                                  |                            |
|------------------|--------------------|----------------------------------|----------------------------|
| <b>Accuracy:</b> | Water tight test:  | Inspection level = G1;           | AQL 1.5                    |
|                  | Visual inspection: | Inspection level = G1;           | AQL 1.5 (critical defects) |
|                  |                    |                                  | AQL 2.5 (major defects)    |
|                  |                    |                                  | AQL 4.0 (minor defects)    |
|                  | Dimensional:       | Inspection level = S2;           | AQL 4.0                    |
|                  | Physical Property: | Inspection level = S2;           | AQL 4.0                    |
|                  | Powder Content:    | Inspection level = 5 pieces;     | Average ≤ 2 mg/glove       |
|                  | Moisture Content:  | Inspection level = 5 pieces;     | Average ≤ 1%               |
|                  | Glove count:       | Inspection level = 5 dispensers; | Tolerance 100+/-2 pieces   |

### Product Specifications:

|                            |                                                               |
|----------------------------|---------------------------------------------------------------|
| <b>Material:</b>           | Nitrile Butadiene Rubber, Not made with natural rubber latex. |
| <b>Sterility</b>           | N/A                                                           |
| <b>Non- Pyrogenic:</b>     | N/A                                                           |
| <b>RNase/ DNase Free:</b>  | N/A                                                           |
| <b>Cytotoxicity:</b>       | N/A                                                           |
| <b>BSE/ TSE Statement:</b> | N/A                                                           |

**Signed :**



Ken Crossley  
Manager  
Quality Assurance

**Date :** June 30, 2015