

CERTIFICATE OF CONFORMANCE

BD Caribe LTD. Junc	os (Subsidiary of Becton Dickinson & Co.) hereby certifies, for the product described below:
Legal Manufacturer /	Address: Becton Dickinson
B-D Catalog No.:	364902
Item Description:	BD LUER-LOK ACCESS DEVICE
Manufacturing Date:	09-2015
Expiration Date:	08-2018

That:

B-D Lot Number:

Quantity: 374,200

5259941

1. The identification of materials used in the manufacturing of this lot of products corresponds to those described below and conform to the requirements as specified in our applicable material specifications:

COMPONENT USED	DESCRIPTION	MATERIAL IDENTIFICATION & FORMULATION	
Male Luer Hub	N/A	Polycarbonate- MAKROLON 2458-550115	
UV Cure Acrylic Adhesive	N/A	Loctite, Product # 3942, Light Cure Acrylic Adhesive	
NP Cannula	N/A	AISI 304 Stainlees Steel	
NP Lubricant	N/A	DC 360 Silicone Oil (12,500 cst)	
Vacutainer	N/A	Basell Polypropylene	
One Use		Resin PF511	
Nonstackable Holder			
NP Sleeve	N/A	Synthetic Isoprene / Synthetic Polyisoprene	

2. A final audit of this product was performed and found acceptable	e as per applicable audit procedure:	VT50123
---	--------------------------------------	---------

Sterilization of this product was performed and found acceptable as per applicable procedure: VS50074
Sterilization Method and Facility Address are described below:

STERILIZATION METHOD	FACILITY (ADDRESS)	STERILIZATION CERTIFICATE
ETHYLENE OXIDE	Steris 2072 Southport Road Spartnburg SC, 29306-6256	Provided by BD Juncos

QA Representative

Date

12-21-15