



Quality Certificate

VWR® NITRILE EXAMINATION GLOVE

As per the manufacturer, the below product meets the following criteria:

| | |
|------------------------------------|---|
| North American Catalog No : | 82026-426 |
| Lot No : | PT234R15J203 |
| Description : | VWR GLOVE NITRILE PF M PK100. |
| Date of Manufacturing : | July 2015 |
| Product Shelf Life : | June 2020 |
| Country of Origin : | Indonesia |
| Quality System Compliance : | FDA 510K, ISO 9001, ISO 13485, CMDCAS Quality System, GMP Certificate |

QC Testing: Representative production samples are collected and inspected in accordance with current applicable Product specification:

- Meet or exceed ASTM D 6319-00a (Water test & Dimension)
- Meet or exceed EN 455 -1&2
- ISO 2859 (Sampling table)

Accuracy:

| | | |
|--------------------|----------------------------------|----------------------------|
| Water tight test: | Inspection level = G1; | AQL 1.5 |
| Visual inspection: | Inspection level = G1; | AQL 1.5 (critical defects) |
| | | AQL 2.5 (major defects) |
| | | AQL 4.0 (minor defects) |
| Dimensional: | Inspection level = S2; | AQL 4.0 |
| Physical Property: | Inspection level = S2; | AQL 4.0 |
| Powder Content: | Inspection level = 5 pieces; | Average ≤ 2 mg/glove |
| Moisture Content: | Inspection level = 5 pieces; | Average ≤ 1% |
| Glove count: | Inspection level = 5 dispensers; | Tolerance 100+/-2 pieces |

Product Specifications:

Material: Nitrile Butadiene Rubber, Not made with natural rubber latex.
Sterility: N/A
Non- Pyrogenic: N/A
RNase/ DNase Free: N/A
Cytotoxicity: N/A
BSE/ TSE Statement: N/A

Signed :



Ken Crossley
 Manager
 Quality Assurance

Date : September 05, 2015