

Greiner Bio-One North America, Inc.
4238 Capital Drive
Monroe NC, 28110, USA



Certificate of Compliance

Product description: VACUETTE® Blood Collection Tube
4 ml K3E K₃EDTA

Greiner product No.: 454012
Lot No.: B141133T
Expiry date: May – 2016

Sterility: Sterile

1	Visual checks	<input checked="" type="checkbox"/>	passed
2	Physical tests	<input checked="" type="checkbox"/>	passed
3	Chemical tests	<input checked="" type="checkbox"/>	passed

We hereby certify that the tubes were tested in accordance with the control procedures stated in the following international standards and meet the required specifications:

EN 14820 "Single-use containers for venous blood specimen collection"
ISO 11137 "Sterilization of health care products – Requirements for validation and routine control – Radiation sterilization"

The product has been manufactured under controlled conditions in accordance with ISO 9001 and EN ISO 13485.

Apr 12, 2015
Date



[Signature] on behalf of
Carolyn Guthrie
QA / QC Manager