

VWR® DISPOSABLE SEROLOGICAL PIPET

As per the manufacturer, the below product meets the following criteria:

| | |
|----------------------------------|-------------------------------|
| North American Catalog No | 89130-894 |
| EU Article No: | 612-4949 |
| Lot Number | 11416007 |
| Description | VWR PIPETTE SERO 2ML PR CS500 |
| Manufacturing Date: | 4/23/2016 |
| Expiration Date: | 4/23/2019 |

Component Materials:

Pipette Polystyrene, meets USP Cytotoxicity Test <87>.
 Mouthpiece Plug Polyester, meets USP Cytotoxicity Test <87>.
 Ink meets USP Cytotoxicity Test <87>.

Volumetric Accuracy: This product has been tested and is accurate within +/- 2% at stated full volume.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

| <u>Test/Procedure</u> | <u>Results</u> |
|-----------------------|----------------|
| Visual Attributes | Pass |
| Volumetric Accuracy | Pass |
| Packaging | Pass |

Sterilization: The lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 "Sterilization of health care products-Requirements for validation and routine control-Radiation sterilization". **Sterility Assurance Level:** SAL 10⁻³

Pyrogens: The product has been tested and has met the criteria established in ANSI/AAMI ST 72:2002/(R) 2010: "Bacterial Endotoxins - Test methodologies, routine testing, and alternative to batch testing". **Results:** <= 0.1 EU/mL (<= 4 EU/device)

Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy: This product is deemed animal free by virtue of not containing materials of animal origin and/or complies with the latest revision of EMA/410/01 section.

Signed:



Ken Crossley
 Manager
 Quality Assurance

Date: May 04, 2016