

02

Certificate of Analysis

2

Becton Dickinson and Company BD Diagnostic Systems PO Box 999 Sparks MD 21152-0999 US

Sparks MD 21152-0999 US Page: 1 of

Product Name : Bottle Starch Agar 500G

Catalog Number : 272100 Manufacture Date :2016/01/20

Batch Number : 6036525 **Expiration Date** : 2020/12/31

01 Dehydrated Medium Appearance: Light beige, free-flowing,

Solubility: 2.5% solution, soluble in distilled or deionized

water on boiling
03 Solution Appearance: Light amber, slightly opalescent

04 Cultural Response: Medium was prepared per label

instructions. Plates were inoculated with a single streak of undiluted test organisms and incubated at 35 ± 2°C for 40-48 hours. Starch hydrolysis was tested by flooding plates with Gram Iodine.

TEST ORGANISMS **ATCC®** RECOVERY STARCH HYDROLYSIS Bacillus subtilis 6633 good positive Escherichia coli 25922 aood negative Staphylococcus aureus 25923 good negative Streptococcus pyogenes 19615 good negative

Characteristic	Unit	Value	Lower Limit	Upper Limit
pH at 25°C : Bulk Lot Number:	_	7.5 6015906	7.3	7.7

Animal source	Country of	Tissue Category			
	Origin	BIC	SIC	ABC	
Bovine	Australia	IV	IV	IB	
Bovine	New Zealand	IV	IV	MLK	
Bovine	USA	III	III	IB	
Bovine	USA	III	IV	IC	
Bovine	USA	IV	IV	IB	
Bovine	USA	IV	IV	NDF	
Porcine	Canada	IV	IV	IB	
Porcine	USA	II	II	IB	
Porcine	USA	III	III	IB	
Porcine	USA	IV	IV	IB	

The Batch Number on this certificate is synonymous with the Lot Number shown on the product label.

BD Diagnostics - Diagnostic Systems products are manufactured in ISO 9001:2008 Registered facilities. In addition, BD Diagnostics - Diagnostic Systems facilities are registered with the United States Food and Drug Administration (FDA), are regulated by the FDA's Quality System Regulations (QSRs), and are also ISO 13485:2003 Registered.

Creation Date: 2016/02/25 16:32:46



Certificate of Analysis

2

Becton Dickinson and Company
BD Diagnostic Systems
PO Box 999

Sparks MD 21152-0999 US Page: 2 of

Product Name : Bottle Starch Agar 500G

Catalog Number : 272100 Manufacture Date :2016/01/20

Batch Number : 6036525
Expiration Date : 2020/12/31

This product met BDDS stringent quality standards at time of batch/lot release. Any test results reported on this certificate were obtained at time of release. This material is not for human or animal consumption.

BD Diagnostics - Diagnostic Systems' Certificates of Analysis (COA) typically contain animal origin information when products are manufactured using materials of animal origin. This information may be contained in the animal source table and/or in one or more of the additional paragraphs found on the COA. Following Quality Control release, the COA is created and published at http://www.bd.com/regdocs. For each batch of finished product that contains animal origin raw materials, the COA shows the animal origin data from the individual lots of animal origin raw materials used, as provided by the raw material suppliers.

At times, suppliers notify BD Diagnostics - Diagnostic Systems of new and/or additional information they have received from their raw material suppliers that modifies the animal origin information for lots previously provided to BD. See "COA Animal Origin Information Position Statement" located at http://www.bd.com/regdocs under "Position Statements" for the impact that retrospective information has on COAs and on customers enrolled in the BDDS and BDAB Automated Change Notification Program.

For complete details on animal origin information, refer to "BD Position Statement - BD Diagnostic-Diagnostic Systems, COA Animal Origin Information Position Statement", at http://www.bd.com/regdocs under "Position Statements".

Manufacturer is Becton Dickinson and Company, 7 Loveton Circle, Sparks, MD 21152 USA. To determine location of manufacturing for this product, please see www.bd.com/ds/technicalCenter/regulatory.asp.

Charlotte Dannenfelser

BD Life Sciences - Diagnostic Systems

Charlette Danserfler

Quality Director, Microbiology

Signature Date:2016/02/25

Creation Date: 2016/02/25 16:32:46