

Date of Certificate: 07/22/2016
Product Number: 5000-0020
Product Lot Number: 1177995
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
1746521	06/16/2016	156	CS	20.2	25.3
1751157	06/20/2016	156	CS	20.7	25.5
1750000	06/21/2016	156	CS	20.3	26.1
1750070	06/22/2016	156	CS	20.0	25.2



Barry Srolis
Sr. Quality Manager