



## Quality Certificate

### VWR® NITRILE EXAMINATION GLOVE

As per the manufacturer, the below product meets the following criteria:

|                                    |   |
|------------------------------------|---|
| <b>North American Catalog No :</b> | 82026-423   |
| <b>Lot No :</b>                    | PT16A001  |
| <b>Description :</b>               | VWR GLOVE NITRILE PF XS PK100   |
| <b>Date of Manufacturing :</b>     | January 2016  |
| <b>Product Shelf Life :</b>        | December 2020   |
| <b>Country of Origin :</b>         | Indonesia   |
| <b>Quality System Compliance :</b> | FDA 510K, ISO 9001, ISO 13485, CMDCAS Quality System, GMP Certificate |

**QC Testing:** Representative production samples are collected and inspected in accordance with current applicable Product specification:

- Meet or exceed ASTM D 6319-00a (Water test & Dimension)
- Meet or exceed EN 455 -1&2
- ISO 2859 (Sampling table)

|                  |                    |                                  |  |
|------------------|--------------------|----------------------------------|--|
| <b>Accuracy:</b> | Water tight test:  | Inspection level = G1;           | AQL 1.5  |
|                  | Visual inspection: | Inspection level = G1;           | AQL 1.5 (critical defects)<br>AQL 2.5 (major defects)<br>AQL 4.0 (minor defects) |
|                  | Dimensional:       | Inspection level = S2;           | AQL 4.0  |
|                  | Physical Property: | Inspection level = S2;           | AQL 4.0  |
|                  | Powder Content:    | Inspection level = 5 pieces;     | Average ≤ 2 mg/glove   |
|                  | Moisture Content:  | Inspection level = 5 pieces;     | Average ≤ 1%   |
|                  | Glove count:       | Inspection level = 5 dispensers; | Tolerance 100+/-2 pieces   |

#### Product Specifications:

|                            |   |
|----------------------------|---|
| <b>Material:</b>           | Nitrile Butadiene Rubber, Not made with natural rubber latex. |
| <b>Sterility</b>           | N/A   |
| <b>Non- Pyrogenic:</b>     | N/A   |
| <b>RNase/ DNase Free:</b>  | N/A   |
| <b>Cytotoxicity:</b>       | N/A   |
| <b>BSE/ TSE Statement:</b> | N/A   |

**Signed :**



Ken Crossley  
Manager  
Quality Assurance

**Date :** 14 January 2016