



Quality Certificate

VWR® NITRILE EXAMINATION GLOVE

As per the manufacturer, the below product meets the following criteria:

| | |
|------------------------------------|---|
| North American Catalog No : | 82026-424 |
| Lot No : | PT16A103 |
| Description : | VWR GLOVE NITRILE PF S PK100 |
| Date of Manufacturing : | Jan 2016 |
| Product Shelf Life : | Dec 2020 |
| Country of Origin : | Indonesia |
| Quality System Compliance : | FDA 510K, ISO 9001, ISO 13485, CMDCAS Quality System, GMP Certificate |

QC Testing: Representative production samples are collected and inspected in accordance with current applicable Product specification:

- Meet or exceed ASTM D 6319-00a (Water test & Dimension)
- Meet or exceed EN 455 -1&2
- ISO 2859 (Sampling table)

| | | | |
|------------------|--------------------|----------------------------------|--|
| Accuracy: | Water tight test: | Inspection level = G1; | AQL 1.5 |
| | Visual inspection: | Inspection level = G1; | AQL 1.5 (critical defects) AQL 2.5 (major defects) AQL 4.0 (minor defects) |
| | Dimensional: | Inspection level = S2; | AQL 4.0 |
| | Physical Property: | Inspection level = S2; | AQL 4.0 |
| | Powder Content: | Inspection level = 5 pieces; | Average ≤ 2 mg/glove |
| | Moisture Content: | Inspection level = 5 pieces; | Average ≤ 1% |
| | Glove count: | Inspection level = 5 dispensers; | Tolerance 100+/-2 pieces |

Product Specifications:

| | |
|-----------------------------|--|
| Material: | Nitrile Butadiene Rubber, Not made with natural rubber latex |
| Sterility | N/A |
| Non- Pyrogenic: | N/A |
| RNase/ DNase Free: | N/A |
| Cytotoxicity : | N/A |
| BSE/ TSE Statement : | N/A |

Signed :



Ken Crossley
Manager
Quality Assurance

Date : Jan 16, 2016