



## Quality Certificate

### VWR® POWDER-FREE NITRILE EXAMINATION GLOVES

As per the manufacturer, the below product meets the following criteria:

<b>North American Catalog No :</b>	82026-430
<b>Lot No :</b>	PT16C008
<b>Description :</b>	VWR GLOVE NITRILE PF XL PK100
<b>Date of Manufacturing :</b>	March 2016
<b>Product Shelf Life :</b>	February 2021
<b>Country of Origin :</b>	Indonesia
<b>Quality System Compliance :</b>	FDA 510K, ISO 9001, ISO 13485, CMDCAS Quality System, GMP Certificate

**QC Testing:** Representative production samples are collected and inspected in accordance with current applicable Product specification:

- Meet or exceed ASTM D3578
- ISO 2859 (Sampling table)

<b>Accuracy:</b>	Water tight test:	Inspection level = G1;	AQL 1.5
	Visual inspection:	Inspection level = G1;	AQL 1.5 (critical defects) AQL 2.5 (major defects) AQL 4.0 (minor defects)
	Dimensional:	Inspection level = S2;	AQL 4.0
	Physical Property:	Inspection level = S2;	AQL 4.0
	Powder Content:	Inspection level = 5 pieces;	Average ≤ 2 mg/glove
	Moisture Content:	Inspection level = 5 pieces;	Average ≤ 1%
	Glove count:	Inspection level = 5 dispensers;	Tolerance 100+/-2 pieces

#### Product Specifications:

<b>Material:</b>	Nitrile Butadiene Rubber
<b>Sterility</b>	N/A
<b>Non- Pyrogenic:</b>	N/A
<b>RNase/ DNase Free:</b>	N/A
<b>Cytotoxicity:</b>	N/A
<b>BSE/ TSE Statement:</b>	N/A
<b>Latex Statement:</b>	This product is not made with natural rubber latex.

**Signed :**



Ken Crossley  
Manager  
Quality Assurance

**Date :** March 17, 2016