

Porcine

Becton Dickinson BD Diagnostic Sys PO Box 999									
Sparks MD 21152-0	)999 US				Page: 1	of 2	3		
Product Name: Bottle Nutrient Agar 500GCatalog Number: 213000Manufacture Date :2016/07/14Batch Number: 6357602Expiration Date: 2021/06/30									
<ul> <li>01. Dehydrated Medium Appearance: Tan, free-flowing, homogeneous</li> <li>02. Solubility: 2.3% solution, soluble in distilled or deionized water on boiling</li> <li>03. Solution Appearance: Light to medium amber, clear to slightly opalescent</li> <li>04. Cultural Response: Medium was prepared per label instructions. Plates were inoculated with the test organisms and incubated at 35 ± 2°C for 18-48 hours. TEST ORGANISMS</li> </ul>									
Enterococcus fa		19433	good						
Escherichia col		25922	good						
Pseudomonas aer	ruginosa	27853	good	d					
Characteristic	Unit	Value		ver nit	Upper Limit				
pH at 25°C : Bulk Lot Number:		6.9 6166920	6.0	5	7.0				
Animal source	Country of Origin		Ti: BIC	ssue Cates SIC	gory ABC				
Bovine	Australia		IV	IV	IB		—		
Bovine	New Zealan	a	IV	IV	MLK				
Bovine	USA		III	III	IB				
Bovine	USA		III	IV	IC				
Bovine	USA USA		IV	IV	IB				
Bovine Porcine	USA Austria		IV IV	IV IV	NDF IB				
Porcine			IV IV	IV IV	IB				
Porcine	Belgium France		IV III	IV	IB				
Porcine			IV	IV	IB				
Porcine	Germany Ireland		IV IV	IV IV	IB				
Porcine			IV IV	IV	IB				
Porcine	Italy Lithuania		IV IV	IV	IB				
Porcine	Netherland	C	IV IV	IV	IB				
Porcine	Poland	Ð	IV IV	IV IV	IB				
			1 V 117	T V T 7	TD				

The Batch Number on this certificate is synonymous with the Lot Number shown on the product label. BD Diagnostics - Diagnostic Systems products are manufactured in ISO 9001:2008 Registered facilities. In addition, BD Diagnostics -

IV

United Kingdom

IV

IΒ



Becton Dickinson and Company BD Diagnostic Systems PO Box 999 Sparks MD 21152-0999 US

Page: 2 of 2

Product Name	Bottle Nutrient Agar 500G
Catalog Number Batch Number	213000 <b>Manufacture Date</b> :2016/07/14 6357602
Expiration Date	2021/06/30

Diagnostic Systems facilities are registered with the United States Food and Drug Administration (FDA), are regulated by the FDA's Quality System Regulations (QSRs), and are also ISO 13485:2003 Registered. This product met BDDS stringent quality standards at time of batch/lot release. Any test results reported on this certificate were obtained at time of release. This material is not for human or animal consumption. BD Diagnostics - Diagnostic Systems' Certificates of Analysis (COA) typically contain animal origin information when products are manufactured using materials of animal origin. This information may be contained in the animal source table and/or in one or more of the additional paragraphs found on the COA. Following Quality Control release, the COA is created and published at http://www.bd.com/reqdocs. For each batch of finished product that contains animal origin raw materials, the COA shows the animal origin data from the individual lots of animal origin raw materials used, as provided by the raw material suppliers. At times, suppliers notify BD Diagnostics - Diagnostic Systems of new and/or additional information they have received from their raw

and/or additional information they have received from their raw material suppliers that modifies the animal origin information for lots previously provided to BD. See "COA Animal Origin Information Position Statement" located at http://www.bd.com/regdocs under "Position Statements" for the impact that retrospective information has on COAs and on customers enrolled in the BDDS and BDAB Automated Change Notification Program.

For complete details on animal origin information, refer to "BD Position Statement - BD Diagnostic-Diagnostic Systems, COA Animal Origin Information Position Statement", at http://www.bd.com/regdocs under "Position Statements".

Manufacturer is Becton Dickinson and Company, 7 Loveton Circle, Sparks, MD 21152 USA. To determine location of manufacturing for this product, please see www.bd.com/ds/technicalCenter/regulatory.asp.

Charlette Danserfilter

Charlotte Dannenfelser BD Life Sciences - Diagnostic Systems Quality Director, Microbiology Signature Date:2017/02/16