

Corning Incorporated - Life Sciences
Antigua Carretera A Roma KM 3.6
San Nicolas De Los Garza NL 66480 MEX
www.corning.com/lifesciences
Refer to website for regional contact information.

Page: 1 / 1

Product Name : 6140 FNL 100MM	Manufacture Date : 2017-01-27
Catalog Number : 6140-100	
Lot ID : 02717064	
Expiration Date : Unconfirmed Validity	

Quality Management System - Complies with the current version of the ISO 9001 Standard.

Glass Type - Borosilicate, Type I, Class A, ASTM E438 - Standard Specification for Glasses in Laboratory Apparatus.

BSE/TSE - Product complies with the latest revision of EMA/410/01 "Note for Guidance on minimising the risk of transmitting animal spongiform encephalopathy agents via human and veterinary medicinal products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMA/410/01.

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

- Visual Inspection - Pass
- Packaging Inspection - Pass
- Dimensional Test - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.



Racelis Acosta Reynoso
Quality Manager