

Date of Certificate: 05/15/2017
Product Number: 5000-0020
Product Lot Number: 1201144
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
1952522	05/09/2017	156	CS	19.8	25.7
1952490	05/09/2017	156	CS	19.7	25.0



Barry Srolis
Sr. Quality Manager