

Date of Certificate: 06/23/2017
Product Number: 5000-0020
Product Lot Number: 1203298
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
1968396	05/30/2017	312	CS	19.8	26.5



Barry Srolis
Sr. Quality Manager