

Corning Incorporated - Life Sciences
 Antigua Carretera A Roma KM 3.6
 San Nicolas De Los Garza NL 66480 MEX
www.corning.com/lifesciences
 Refer to website for regional contact information.

Page: 1 / 1

Product Name	: 9825 RG SCT 13X100 TUBE COMP	W/ SCREW CAP
Catalog Number	: 9825-13	Manufacture Date : 2017-08-23
Lot ID	: 23517009	
Expiration Date	: Unconfirmed Validity	

Quality Management System - Complies with the current version of the ISO 9001 Standard.

Glass Type - Borosilicate, Type I, Class A, ASTM E438 - Standard Specification for Glasses in Laboratory Apparatus.

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass
 Packaging Inspection - Pass
 Dimensional Test - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.



Miguel Flores
 Quality Manager