

Date of Certificate: 12/06/2017
Product Number: 5000-0020
Product Lot Number: 1215868
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
2094079	11/29/2017	156	CS	20.9	26.4
2098928	12/04/2017	26	CS	19.6	26.5
2099866	12/05/2017	130	CS	20.2	26.5



Barry Srolis
Sr. Quality Manager