


---

Date of Certificate: 11/03/2017  
Product Number: 5000-0020  
Product Lot Number: 1212887  
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
2069899	10/25/2017	60	CS	20.6	26.2
2075351	10/30/2017	156	CS	20.7	26.8
2075095	11/02/2017	96	CS	21.1	26.7



---

**Barry Srolis**  
Sr. Quality Manager