

Date of Certificate: 12/06/2017
Product Number: 5000-0020
Product Lot Number: 1216421
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
2098928	12/04/2017	156	CS	19.6	26.5
2099866	12/05/2017	156	CS	20.2	26.5



Barry Srolis
Sr. Quality Manager