

Date of Certificate: 01/16/2018
Product Number: 5000-0020
Product Lot Number: 1217682
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
2114458	12/27/2017	156	CS	19.8	25.6
2120845	01/08/2018	17	CS	19.8	25.5
2120916	01/09/2018	82	CS	19.5	26.3
2125492	01/16/2018	57	CS	20.3	26.2



Barry Srolis
Sr. Quality Manager