

Certificate Of Analysis

Whatman™

Product Information

Product Number: 3001-604

Product Name: GR 1CHR 1CMx100M

Lot Number/Certificate No: 968382801

Parent Reel Job No: 9673554

Manufacture Date of Parent Reel: 5 AUG 2015

Reel Number: RL1520004

Quality Control

Test	Range	Results		
		Mean	Min	Max
Grammage,gsm	80-93	90	89	90
Thickness, $\mu\text{m}@53\text{Kpa}$	160-196	185	179	189
Porosity,s/200ml/sq=in	13-31	22	21	22

Conformance & Quality systems statement

This is to certify that this product conforms to GE Healthcare Life Sciences specifications.

All products are manufactured via a Quality System certified to **BS EN ISO9001:2008** and tested in accordance with documented quality procedures and approved as a result of meeting the required specification

Electronic signature

This document has been electronically produced and is valid without a signature.

Version AC