


---

Date of Certificate: 02/21/2018  
Product Number: 5000-0020  
Product Lot Number: 1221390  
Specified Dose Range: 19.0 - 28.0 KGY

| Run ID  | Receipt Date | Quantity | UOM | Delivered Min | Delivered Max |
|---------|--------------|----------|-----|---------------|---------------|
| 2149833 | 02/20/2018   | 313      | CS  | 20.5          | 26.5          |



---

**Barry Srolis**  
Sr. Quality Manager