

Date of Certificate: 05/23/2018
Product Number: 5000-0020
Product Lot Number: 1226387
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
2203213	05/15/2018	156	CS	20.9	26.6
2203268	05/15/2018	110	CS	20.7	26.7
2207953	05/23/2018	46	CS	20.7	26.3



Barry Srolis
Sr. Quality Manager