



## VWR® LABORATORY BOTTLES, HIGH-DENSITY POLYETHYLENE

As per the manufacturer, the below product meets the following criteria:

<b>VWR Catalog Number</b>	414004-159
<b>VWR Europe Catalog Number</b>	215-3408
<b>Lot Number</b>	VHF140418
<b>Description</b>	VWR BOTTLE HDPE NM 1000ML PK6
<b>Date of Manufacture</b>	April 14, 2018
<b>Country of Origin</b>	India

### Quality System Compliance

Products are manufactured, inspected and found to be in compliance with product and quality specification requirements as documented in the ISO 9001:2008 & ISO13485:2003 Quality Management System. Inspection records are reviewed and signed for product release.

### QC Testing

Representative production samples are collected and inspected in accordance with current applicable standard protocols.

Test Name	Test Reference	Accuracy	Result
Leak Test	QWI-QC-03	General Insp. Level - I, AQL - 1.5	Leak proof
Autoclaving	QWI-QC-07	General Insp. Level - I, AQL - 1.5	NA
Dimensional	QSF-QC-02	General Insp. Level - I, AQL - 1.5	Within tolerance

### Product Specifications

**Material:** **Bottle:** High-Density Polyethylene, 100 % Virgin, FDA 21 CFR Compliance, USP Class VI passed.

**Closure:** Polypropylene, 100 % Virgin, FDA 21 CFR Compliance, USP Class VI passed.

**BSE/TSE Statement:** The above-mentioned product is manufactured without the use of Raw materials of animal origin. Therefore, BSE/TSE issue does not concern this product to the best of the manufacturer's knowledge.

**Non-Pyrogenic:** Not Applicable.

**DNase & RNase Free:** Not Applicable.

**Cytotoxicity:** Not Applicable.

**Latex-Free Statement:** No materials containing latex or natural rubber are used in the manufacturing, handling and packaging processes for this product.

Signed:



Linda K. Shine  
 Manager - Supplier Qualification  
 Global Compliance  
 VWR International LLC

**Date:** April 14, 2018