

Date of Certificate: 05/31/2018
Product Number: 5000-0020
Product Lot Number: 1229702
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
2212655	05/29/2018	156	CS	21.4	26.9
2211772	05/30/2018	154	CS	20.5	26.4



Barry Srolis
Sr. Quality Manager