



VWR® LABORATORY BOTTLES, HIGH-DENSITY POLYETHYLENE

As per the manufacturer, the below product meets the following criteria:

VWR Catalog Number	414004-159
VWR Europe Catalog Number	215-3408
Lot Number	VHF160418
Description	VWR BOTTLE HDPE NM 1000ML PK6
Date of Manufacture	April 16, 2018
Country of Origin	India

Quality System Compliance

Products are manufactured, inspected and found to be in compliance with product and quality specification requirements as documented in the ISO 9001:2008 & ISO13485:2003 Quality Management System. Inspection records are reviewed and signed for product release.

QC Testing

Representative production samples are collected and inspected in accordance with current applicable standard protocols.

Test Name	Test Reference	Accuracy	Result
Leak Test	QWI-QC-03	General Insp. Level - I, AQL - 1.5	Leak proof
Autoclaving	QWI-QC-07	General Insp. Level - I, AQL - 1.5	NA
Dimensional	QSF-QC-02	General Insp. Level - I, AQL - 1.5	Within tolerance

Product Specifications

- Material:** *Bottle:* High Density Polyethylene, 100 % Virgin, FDA 21 CFR Compliance, USP Class VI passed.
- Closure:* Polypropylene, 100 % Virgin, FDA 21 CFR Compliance, USP Class VI passed.
- BSE/TSE Statement:** The above-mentioned product is manufactured without the use of Raw materials of animal origin. Therefore, BSE/TSE issue does not concern this product to the best of the manufacturer's knowledge.
- Non-Pyrogenic:** Not Applicable.
- DNase & RNase Free:** Not Applicable.
- Cytotoxicity:** Not Applicable.
- Latex-Free Statement:** No materials containing latex or natural rubber are used in the manufacturing, handling and packaging processes for this product.

This document has been produced electronically and is valid without a signature.

Signed on behalf of VWR International:

Michael J. Mascali, RPh, FASCP, ASQ CMQ – OE
 SVP Compliance
 VWR International LLC

Date: April 16, 2018