


Certificate of Quality



As per the manufacturer, the below product meets the following criteria:

To per une meneral and a sole in product more than sole une sole u	
VWR Catalog Number	83007-380
European Article Number	613-0342
Lot Number	60423-744C6-744A
Description	VWR TIP BLU 1000UL RK ST PK576
Date of Manufacture	October 30, 2017
Date of Sterilization	November 06, 2017
Date of Expiration	2020-10
Country of Origin	Made in USA

Quality System Compliance

The product conforms to written material specifications and was manufactured under the manufacturer's registered and audited ISO 9001 quality system, and underwent lot testing as outlined in the manufacturer's laboratory procedures. These products come with the highest standard of quality assurance.

QC Testing

Each lot of item was produced in a tightly controlled environment and subjected to the manufacturer's rigorous testing and performance procedures. The products meet all stated standards for precision, clarity, warp, centrifugation and freedom from contamination.

Product Specifications

Material: Pipet Tip Material: Polypropylene

.

Certified free of Bisphenol A (BPA), phthalates and cytotoxic effects. Plastic resins used in product manufacturing have been tested for heavy metals using the prescribed USP method and confirmed to have levels lower than 1 PPM. Resins are USP Class-VI certified, RoHS, FDA regulation CFR 21, and are free of Substances of Very High Concern (SVHC), REACH

compliant, and are FDA approved for food contact.

Non-Pyrogenic Statement: Product samples are exposed to endotoxin-free water and the resulting extraction fluid is tested

for contamination using the kinetic turbidimetric Limulus Amoebocyte Lysate (LAL) assay protocol and USP guidelines. All products tested must display less than 0.05 EU/ml to be

certified free of endotoxin.

ATP Assay: Product sample surfaces are tested for the presence of Adenosine triphosphate (ATP) using

a controlled bioluminescence reaction to detect contamination. Luminescence data is compared to results generated by ATP-free surfaces and surfaces with known amounts of ATP as a positive control. The relative light units' result must indicate less than 2X10⁻¹² mg/ul

of ATP for the product to be certified as ATP free.

DNase & RNase Free: Product samples are exposed to nuclease-free water and the resulting extraction fluid is tested

for nuclease activity on commercially available 7.5 kb Poly(A) tailed RNA (1ug) and HindIII-digested DNA (1ug) with a one hour 37°C incubation in appropriate buffers. Results are visualized on an agarose gel with appropriate positive and negative controls. Extraction fluid

VWR International LLC, Radnor Corporate Center, Building One, Suite 200, 100 Matsonford Road Radnor, PA 19087 VWR International bvba/sprl, Haasrode Research Park Zone 2020, Geldenaaksebaan 464, 3001 Leuven, Belgium http://www.vwr.com

Page 1 of 2 12145



Certificate of Quality



samples must show no degradation of the nucleic acids by the extraction fluid has occurred for

the product to be certified as RNase-free and DNAse-free.

BSE/TSE Statement: (Bovine Spongiform Encephalopathy/Transmissible Spongiform Encephalopathy)

The product above is free from and manufactured from materials that do not contain any raw materials or substances derived from animal origin as defined in EC Directive 97/534/EC and

EC Directive 76/768/EEC and Amendment 419 Annex II.

The manufacturer does not store any products of animal origin, including animal proteins, in

any manufacturing areas of their facility or warehouses.

Sterilization Process: Products from the specified lot number was processed by e-beam radiation at a dose range of

15.2-32.0 kiloGray. This dosage is sufficient to guarantee a sterility assurance level of 10⁻⁶. Sterility assurance levels are based on the probability of a positive, nonsterile part occurring after irradiation. An SAL of 10⁻⁶ is the highest level of sterility able to be guaranteed according to the ISO 11137 standard and represents a 1 in 1,000,000 chance of a nonsterile part occurring. The Manufacturer's sterilization program is an ISO 11137 validated process, with bioburden studies

and radiation validation audits performed quarterly.

Specified Dosage Range: 15.20 - 32.00 kGy

Minimum Dosage Delivered: 15.7 kGy

Maximum Dosage Delivered: 29.6 kGy

Latex Statement: No Latex was used in the manufacturing of the product. This also includes all the packaging and

shipping materials used in the production of all items.

The only time the product or packaging may come in contact with Latex would be from the Latex gloves that production employees are required to wear for the purpose of clean handling and to

avoid biological contamination of the products.

Disclaimer: VWR states that this declaration will not discharge the user from his obligation to ensure the product is suitable for the intended use.

This Certificate was automatically generated and is valid without a Signature.

Date: June 27, 2018