

Corning Incorporated - Life Sciences  
271 County Route 64  
Big Flats NY 14814 USA  
www.corning.com/lifesciences  
Refer to website for regional contact information.

Page: 1 / 1

<b>Product Name</b>	: BOTTLE MEDIA SQ 500ML CAP/RING 21820442		
<b>Catalog Number</b>	: 1396-500	<b>Manufacture Date</b>	: 2018-03-29
<b>Lot ID</b>	: 08818001		
<b>Expiration Date</b>	: Unconfirmed Validity		

**Quality Management System** - Complies with the current version of the ISO 9001 Standard.

**Glass Type** - Borosilicate, Type I, Class A, ASTM E438 - Standard Specification for Glasses in Laboratory Apparatus.

**BSE/TSE** - Product complies with the latest revision of EMA/410/01 "Note for Guidance on minimising the risk of transmitting animal spongiform encephalopathy agents via human and veterinary medicinal products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMA/410/01.

**Quality Control Testing** - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

- Visual Inspection - Pass
- Packaging Inspection - Pass
- Dimensional Test - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.



Elizabeth M (Betty) Hansen  
Quality Manager