

Date of Certificate: 01/23/2018
Product Number: 5000-0020
Product Lot Number: 1219032
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
2125433	01/16/2018	156	CS	20.5	26.2
2126287	01/17/2018	116	CS	19.1	24.5



Barry Srolis
Sr. Quality Manager