


---

Date of Certificate: 03/29/2018  
Product Number: 5000-0020  
Product Lot Number: 1224721  
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
2167743	03/21/2018	218	CS	20.4	26.0
2165216	03/28/2018	94	CS	21.6	26.3



---

**Barry Srolis**  
Sr. Quality Manager