

Date of Certificate: 06/28/2018
Product Number: 5000-0020
Product Lot Number: 1231172
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
2221864	06/26/2018	312	CS	20.1	26.5



Barry Srolis
Sr. Quality Manager