

Date of Certificate: 08/16/2018
Product Number: 5000-0020
Product Lot Number: 1233390
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
2263144	08/14/2018	175	CS	20.3	25.7
2262923	08/15/2018	135	CS	20.1	25.9



Barry Srolis
Sr. Quality Manager