

Certificate of Analysis



HistoChoice Tissue Fixative A

H112

Material: Grade:

HISTOLOGY GRADE

Batch Number: 18A2956210

Chemical Formula:	Not Applicable	Manufacture Date: 07/19/201		07/19/2017	
Molecular Weight:	Not applicable	Reassay Date: 07/31/2019		07/31/2019	
CAS #:	Not Applicable				
Appearance:		Storage: Room Temperature			
Colorless to Light Yellow Haze-free Liquid					

TEST	SPECIFICATION	ANALYSIS	DISPOSITION
Abs. @ 200-360nm (10%, Water)	Report	0.5445	PASS
Appearance	PASS / Fail	PASS	PASS
Lambda Max @ 200-360nm (10%, Water)	293 - 299 nm	295 nm	PASS
pH @ 25C	5.0 - 6.5	5.8	PASS
Water Content (KF)	78 - 86 %	82 %	PASS

Spec Set: H112HG

Internal ID #: 587

Signature	Additional Information
We certify that this batch conforms to the specifications listed.	Analysis may have been rounded to significant digits in specification limits.
This document has been electronically produced and is valid without a signature.	Product meets analytical specifications of the grades listed.
Leona Edwardson, Quality Control Sr. Manager - Solon VWR Chemicals, LLC. 28600 Fountain Parkway, Solon OH 44139 USA	