



VWR International, Inc.
1310 Goshen Parkway,
West Chester, PA 19380
610-431-1700, www.vwr.com

CERTIFICATE OF ANALYSIS

VWR International certifies that the below lot was manufactured in compliance with the applicable specifications and regulatory requirements.

Product Number	48218-402
Description	Resolve® Microscope Immersion Oil, High Viscosity
Lot Number	102691
Expiration Date:	03/2010

Test	Specification	Results
Appearance	Clear Liquid	Pass
Viscosity @ 23C	1037.7 – 1074.3	1044cPs
Refractive Index @ 23C	1.5149 – 1.5151	1.5151

Signed:

A handwritten signature in blue ink, appearing to read "Ken Crossley".

Ken Crossley
Manager
Quality Assurance